## PAX-8



## IN VITRO DIAGNOSTIC DATASHEET

## INTENDED USE : IN VITRO DIAGNOSTIC USE

This product is intended for qualitative immunohistochemistry with normal and neoplastic formalin-fixed, paraffin-embedded tissue sections, to be viewed by light microscopy.

## **DESCRIPTION :**

PAX-8 antibody is expressed in the thyroid (and associated carcinomas), non-ciliated mucosal cells of the fallopian tubes and simple ovarian inclusion cysts, but not normal ovarian surface epithelial cells.

| CATALOG NO : PL3786   | PL3786-R7<br>PL3786-R1<br>PL3786-1<br>PL3786-0,1 | 7 ML RTU 70 TEST<br>1 ML RTU 10 TEST<br>1 ML 1/100 1000 TEST<br>0,1 ML 1/100 100 TEST |
|---|--|---|
| STAINING PATTERN : Nuclear  |  |   |
| POSITIVE CONTROL : Ovary, Thyroid   | ŗ  |   |
| VOLUME : 7 ml Ready to Use ( 7 ml of antibody prediluted in 0.05mol/L Tris-HCl, pH 7.6 containing stabilizing protein and 0.015mol/L sodium azide. )   HOST : Mouse   |  |   |
|   |  |   |
| ANTIBODY CONCENTRATION : 200ug / ml   |  |   |
| SPECIES REACTIVITY: Human. Others not tested.   |  |   |
| EPITOPE : N/A   |  |   |
| MICROBIOLOGICAL STATE : This product is not sterile.  |  |   |
| <b>PRETREATMENT :</b> Staining of formalin-fixed tissue sections requires treating the tissue sections in boiling 10mM citrate buffer, pH 6.0, for 10-20 minutes followed by cooling at room temperature for 20 min.  |  |   |
| PRIMARY ANTIBODY INCUBATION TIME : 30 minutes at Room Temperature   |  |   |
| <b>STAINING TIPS :</b> If the staining is too light, use lower dilution or longer time. If the staining is too strong, check pretreatment, use higher dilution or shorter time.   |  |   |
| <b>STORAGE AND STABILITY :</b> This product contains sodium azide and is stable for 24 months when stored at 2-8°C. Do not use after expiration date indicated on label of the product. If reagent is not stored as recommended, performance must be validated by the user. |  |   |
| <b>TROUBLESHOOTING :</b> Please contact Patolab Technical Support by e-mail ( patolab@patolab.com.tr ).   |  |   |

